

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p style="text-align: center;"><b>Accredited to ISO/IEC 17025:1999</b></p>	<b>Euro Lab Environmental Ltd</b>	
	<b>Issue No: 005</b>	<b>Issue date: 27 November 2006</b>
	<b>Units G1 and G2</b> Peartree House 1 Britannia Road Brentwood Essex CM14 5LD	<b>Contact: Mr E Taylor</b> Tel: +44 (0)1277 210022 Fax: +44 (0)1277 233049 E-Mail: eurolab@btconnect.com Website: www.eurolab.co.uk
<b>Testing performed by the Organisation at the locations specified below</b>		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code		
<table style="width: 100%; border: none;"> <tr> <td style="width: 50%; border: none;"> <b>Address</b>                      Units G1 and G2                      Peartree House                      1 Britannia Road                      Brentwood                      Essex                      CM14 5LD                 </td> <td style="width: 50%; border: none;"> <b>Local contact</b>                      Mr E Taylor                       Tel: +44 90)1277 210022                      Fax: +44 1277 233049                      Email: eurolab@btconnect.com                      Website: www.eurolab.co.uk                 </td> </tr> </table>	<b>Address</b> Units G1 and G2 Peartree House 1 Britannia Road Brentwood Essex CM14 5LD	<b>Local contact</b> Mr E Taylor  Tel: +44 90)1277 210022 Fax: +44 1277 233049 Email: eurolab@btconnect.com Website: www.eurolab.co.uk	Health and Hygiene	A
<b>Address</b> Units G1 and G2 Peartree House 1 Britannia Road Brentwood Essex CM14 5LD	<b>Local contact</b> Mr E Taylor  Tel: +44 90)1277 210022 Fax: +44 1277 233049 Email: eurolab@btconnect.com Website: www.eurolab.co.uk			

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
Premises including domestic, commercial and industrial	Health and Hygiene	B



2664  
Accredited to  
ISO/IEC 17025:1999

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Euro Lab Environmental Ltd**  
**Issue No: 005 Issue date: 27 November 2006**

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248) and Methods for the Determination of Hazardous Substances in the series MDHS	
	Fibre counting	HSG 248:February 2005 (Documented In-House Procedure)  MDHS 39/4:November 1995 Membrane Filter Method using Phase Contrast Microscopy (PCM)	A & B
	Sampling of air for fibre counting	HSG 248:February 2005 (Documented In-House Procedure) in QM 2D  MDHS 39/4:November 1995 (Documented In-House Procedure) in QM 2E	B
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	HSG 248:February 2005 (Documented In-House Procedure)  MDHS 39/4:November 1995 Membrane Filter Method using Phase Contrast Microscopy (PCM)	B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248:February 2005 by Documented In-House Procedure in the Quality Manual at Book 2, Part C, using stereo-microscopy, polarised light microscopy and dispersion staining	A
	Sampling of bulk materials for asbestos identification	Documented In-House Procedure in QM 2C	B

END